

Forte JETStream

Cleaning

You can clean the outside of the Forte imaging system components by doing the following:

- To clean the plastic or metal imaging system components, moisten a soft, static-free cloth with a non-abrasive cleaning agent and wipe the outside of the object.
- To clean the imaging table, use a disinfectant in accordance with your site protocol.

Precautions

When cleaning the imaging system equipment, observe the following precautions:

- Do not spray aerosols or liquids on the imaging system equipment.
- Do not use solvents or abrasives on the imaging system equipment.
- Do not clean the internal components of the imaging system.



CAUTION: If a spill occurs and liquid enters the gantry, imaging table, or acquisition equipment, turn off the equipment, and then call a Philips Nuclear Medicine Field Service Engineer.

Care and Maintenance of the Touchscreen

- To avoid risk of electric shock, do not disassemble the A/C adaptor or display unit cabinet. The unit is not user-serviceable. Remember to unplug the display unit from the power outlet before cleaning.
- Do not use alcohol (methyl, ethyl, or isopropyl) or any strong solvent. Do not use thinner or benzene, abrasive cleaners or compressed air.
- To clean the display unit cabinet, use a cloth lightly dampened with a mild detergent.
- Avoid getting liquids inside your touchscreen. If liquid does get inside, contact your Philips field service engineer before you power it on again.
- Do not wipe the screen with a cloth or sponge that could scratch the surface.
- To clean the touchscreen, use window or glass cleaner. Put the cleaner on the rag and wipe the touchscreen. Never apply the cleaner directly on the touchscreen.